



Report To

Gorky Gonzalez  
 Gorky Pottery - Laredo  
 Alfareria Tradicional SA de CV  
 Ex Huerta de Moontenogonr S/N  
 Col. Pastita Gu, 36090

Project

326586

Quality Control

196594		M	Solid	ASTM C738-81				
<b>Blank</b>								
Parameter	PrepSet	Reading	MDL	MQL	Units		Out	File
Lead (Acetic Acid Extractable)	195987	ND	5.00	0	ug/L			0002890091
<b>CCV</b>								
Parameter		Reading	Known	Units	Recover%	Limits%	Out	File
Lead (Acetic Acid Extractable)		299	300	ug/L	99.7	90.0 - 110		0002890098
Lead (Acetic Acid Extractable)		309	300	ug/L	103	90.0 - 110		0002890089
Lead (Acetic Acid Extractable)		315	300	ug/L	105	90.0 - 110		0002890101
<b>ICV</b>								
Parameter		Reading	Known	Units	Recover%	Limits%	Out	File
Lead (Acetic Acid Extractable)		104	100	ug/L	104	90.0 - 110		0002890088
<b>LCS</b>								
Parameter	PrepSet	Reading	Known	Units	Recover%	Limits	File	Out
Lead (Acetic Acid Extractable)	195987	95.8	100	ug/L	95.8	85.0 - 115	0002890092	

RPD is Relative Percent Difference:  $\text{abs}(r1-r2) / \text{mean}(r1,r2) * 100\%$

Recover% is Recovery Percent:  $\text{result} / \text{known} * 100\%$

